

Certificate of Analysis

PRODUCT : Karl Fischer water standard 1.0 mg/g RS - ERBAqua - For checking calibration of coulometric Karl Fischer equipment

CODE : 570211

LOT N° : 20260097

RE-TEST DATE : 07/2025

TEST	U.M.	SPECIFICATION	RESULT
Water (K.F.)	mg/g	0.97 - 1.03	1.00

It is under the customer responsibility to make sure the purchased product is suitable for his use and/or application.
CARLO ERBA Reagents would not be guilty of any misuse or mishandling of any of its products, occurring potential damages or user hurts in case of inappropriate use.

Approve Date : 04/11/2020

Approved electronically, valid without a signature

QUALITY CONTROL RESPONSIBLE

B. COULANGE (VDR)