

Certificate of Analysis

PRODUCT : Karl Fischer water standard 1.0 mg/g RS - ERBAqua - For checking calibration of coulometric Karl Fischer equipment

CODE : 570211

LOT N° : 20390192

RE-TEST DATE : 12/2025

TEST	U.M.	SPECIFICATION	RESULT
Water (K.F.)	mg/g	0.970 - 1.030	0.996

Traceable to the SRM 2890 of NIST (National Institute of Standards and Technology)

It is under the customer responsibility to make sure the purchased product is suitable for his use and/or application.
 CARLO ERBA Reagents would not be guilty of any misuse or mishandling of any of its products, occurring potential damages or user hurts in case of inappropriate use.

Approve Date : 08/12/2021

Approved electronically, valid without a signature

QUALITY CONTROL RESPONSIBLE

B. COULANGE (VDR)