

# Certificate of Analysis

**PRODUCT :** Karl Fischer water standard 1.0 mg/g RS - ERBAqua - For checking calibration of coulometric Karl Fischer equipment

**CODE :** 570211

**LOT N° :** 21290254

**RE-TEST DATE :** 07/2026

TEST	U.M.	SPECIFICATION	RESULT
Water (K.F.)	mg/g	0.970 - 1.030	1.011

Traceable to N.I.S.T.

It is under the customer responsibility to make sure the purchased product is suitable for his use and/or application.  
 CARLO ERBA Reagents would not be guilty of any misuse or mishandling of any of its products, occurring potential damages or user hurts in case of inappropriate use.

Approve Date : 27/10/2021

Approved electronically, valid without a signature

**QUALITY CONTROL RESPONSIBLE**

**B. COULANGE (VDR)**