

# Certificate of Analysis

**PRODUCT :** Karl Fischer water standard 1.0 mg/g RS - ERBAqua - For checking calibration of coulometric Karl Fischer equipment  
**CODE :** 570211  
**LOT N° :** 24010768  
**RE-TEST DATE :** 04/2029

TEST	U.M.	SPECIFICATION	RESULT
Water (K.F.)	mg/g	0.970 - 1.030	1.007

Traceable to the SRM 2890 of NIST (National Institute of Standards and Technology)

It is under the customer responsibility to make sure the purchased product is suitable for his use and/or application.  
 CARLO ERBA Reagents would not be guilty of any misuse or mishandling of any of its products, occurring potential damages or user hurts in case of inappropriate use.

Approve Date : 22/05/2024

Approved electronically, valid without a signature

**QUALITY CONTROL RESPONSIBLE**  
**B. COULANGE (VDR)**