

# Certificate of Analysis

**PRODUCT** : KARL FISCHER WATER STANDARD 1.0 mg/g RS ERBAqua for calibration  
**CODE** : 570211  
**LOT N°** : C794400  
**EXPIRING DATE** : 2022/02 **EDITION** 1

TEST	U.M.	SPECIFICATION	RESULT
Water (K.F.)	mg/g	0.97 ÷ 1.03	0.99

Traceable to N.I.S.T.