



Carlo Erba Reagents S.A.S.
 Parc d'Affaires des Portes
 Chaussée du Vexin - BP 616
 27106 Val de Reuil Cedex
 Tél. : 02.32.09.20.00
 Fax 02.32.09.20.20

Z.I DE VALDONNE - BP 4
 13124 PEYPIN
 Tél. : 04.42.32.41.41
 Fax 04.42.72.41.62

Carlo Erba Reagents S.r.l.
 Via R. Merendi, 22
 20010 Cornaredo (Mi)
 Tel. 02 93 991 90 Fax 02 93 991 001

ISO 9001: 2015

Technical Specification

PRODUCT : Standard solution 20000 µS/cm RS - For conductivity

CODE : 575140

TEST	U.M.	SPECIFICATION
Description	-	Clear colourless liquid
Conductivity at 25°C	µS/cm	19800 - 20200

Traceable to Standard NIST N° 999c
 Confidence interval ± 100 µS/cm

It is under the customer responsibility to make sure the purchased product is suitable for his use and/or application.
 CARLO ERBA Reagents would not be guilty of any misuse or mishandling of any of its products, occurring potential damages or user hurts in case of in

Date : 01/08/2023

QUALITY CONTROL RESPONSIBLE
B. COULANGE (VDR)