



Carlo Erba Reagents S.A.S.  
Parc d'Affaires des Portes  
Chaussée du Vexin - BP 616  
27106 Val de Reuil Cedex  
Tél.:02.32.09.20.00  
Fax 02.32.09.20.20

Z.I DE VALDONNE - BP 4  
13124 PEYPIN  
Tél.:04.42.32.41.41  
Fax 04.42.72.41.62

Carlo Erba Reagents S.r.l.  
Via R. Merendi, 22  
20010 Cornaredo (Mi)  
Tel. 02 93 991 90 Fax 02 93 991 001

ISO 9001: 2015

# Technical Specification

**PRODUCT : Standard solution 1.30  $\mu\text{S}/\text{cm}$  RS For conductivity**

**CODE : 575230**

TEST	U.M.	SPECIFICATION
Description	-	Clear colourless liquid
Conductivity at 25°C	$\mu\text{S}/\text{cm}$	1.25 - 1.35

Traceable to Standard NIST N° 999a  
Confidence interval:  $\pm 0.02 \mu\text{S}/\text{cm}$

It is under the customer responsibility to make sure the purchased product is suitable for his use and/or application.  
CARLO ERBA Reagents would not be guilty of any misuse or mishandling of any of its products, occurring potential damages or user hurts in case of in

Date : 29/11/2022

**QUALITY CONTROL RESPONSIBLE**  
**B. COULANGE (VDR)**