

Technical Specification

PRODUCT : BUFFER SOLUTION pH 4.5 RS - For HPLC
CODE : 524970

TEST	U.M.	SPECIFICATION
Clear,colourless solution	-	Conform
pH	unite pH	4.45 - 4.55
Temperature of measurement	°C	19 - 21

Accuracy : +/- 0.05 pH units - Standardized against NIST reference materials.
Filtered through 0.45µm membrane

It is under the customer responsibility to make sure the purchased product is suitable for his use and/or application.
CARLO ERBA Reagents would not be guilty of any misuse or mishandling of any of its products, occurring potential damages or user hurts in case of in